

1939

A Deuterium - Deuterium Source of Neutrons

Victor Young
State University of Iowa

Let us know how access to this document benefits you

Copyright ©1939 Iowa Academy of Science, Inc.

Follow this and additional works at: <https://scholarworks.uni.edu/pias>

Recommended Citation

Young, Victor (1939) "A Deuterium - Deuterium Source of Neutrons," *Proceedings of the Iowa Academy of Science*, 46(1), 266-266.

Available at: <https://scholarworks.uni.edu/pias/vol46/iss1/75>


This Research is brought to you for free and open access by the IAS Journals & Newsletters at UNI ScholarWorks. It has been accepted for inclusion in Proceedings of the Iowa Academy of Science by an authorized editor of UNI ScholarWorks. For more information, please contact scholarworks@uni.edu.

Offensive Materials Statement: Materials located in UNI ScholarWorks come from a broad range of sources and time periods. Some of these materials may contain offensive stereotypes, ideas, visuals, or language.

A DEUTERIUM — DEUTERIUM SOURCE OF NEUTRONS


VICTOR YOUNG

A single section accelerator tube built for 300 to 400 k.v. is used to accelerate a beam of deuterons obtained from a conventional arc discharge source. The beam impinges on a target of $P_2O_5 + H_2^2O$ producing neutrons according to the reaction



The target assembly is immersed in a tank of water which because of the elastic H^1 and n collisions becomes a source of thermal neutrons.

Detection is accomplished by placing a piece of silver in the tank. The silver becomes artificially beta radioactive by the well known reactions



Since the half life of the beta activity is something over three minutes there is ample time to remove the silver from the tank and detect the beta particles with a thin walled Geiger-Müller counter.

DEPARTMENT OF PHYSICS,
STATE UNIVERSITY OF IOWA,
IOWA CITY, IOWA.

APPARATUS FOR PRODUCING SOFT X-RAYS

F. M. BAILEY

A 30-watt soft x-ray apparatus has been designed to provide a simple and economical source of soft x-rays for radiography in the wavelength region between 1.1 and 0.6 Angstroms. The x-ray tube was constructed of pyrex, and a thin spherical window incorporated for transmitting the radiation. The electrical equipment was built from standard laboratory parts, employing a neon sign transformer as a source of potential.

DEPARTMENT OF PHYSICS,
IOWA STATE COLLEGE,
AMES, IOWA.