

1936

## Biological Fluid Osmotic Pressure Determinations

E. C. McCracken  
*Iowa State College*

*Let us know how access to this document benefits you*

Copyright ©1936 Iowa Academy of Science, Inc.

Follow this and additional works at: <https://scholarworks.uni.edu/pias>

---

### Recommended Citation

McCracken, E. C. (1936) "Biological Fluid Osmotic Pressure Determinations," *Proceedings of the Iowa Academy of Science*, 43(1), 270-270.

Available at: <https://scholarworks.uni.edu/pias/vol43/iss1/83>

This Research is brought to you for free and open access by the IAS Journals & Newsletters at UNI ScholarWorks. It has been accepted for inclusion in Proceedings of the Iowa Academy of Science by an authorized editor of UNI ScholarWorks. For more information, please contact [scholarworks@uni.edu](mailto:scholarworks@uni.edu).

**Offensive Materials Statement:** Materials located in UNI ScholarWorks come from a broad range of sources and time periods. Some of these materials may contain offensive stereotypes, ideas, visuals, or language.

BIOLOGICAL FLUID OSMOTIC PRESSURE  
DETERMINATIONS

E. C. McCRACKEN

An outgrowth of the method used by A. V. Hill to determine the next liberation accompanying the passage of a nerve impulse has resulted in the development of an apparatus by which the osmotic pressure of as small quantities of fluid as .5 cu. mm. may be determined to an accuracy of .1 per cent NaCl. The principle consists essentially in the differential cooling effect produced between the experimental sample and a standard solution when placed in a standard, constant temperature, constant humidity, environment. A discussion of the necessary thermocouple loop construction, electrical difficulties, and optical system is included in the report.

DEPARTMENT OF PHYSICS,  
IOWA STATE COLLEGE,  
AMES, IOWA.

---

  
MAGNETIC VELOCITY ANALYSIS OF POTASSIUM  
ATOMS SCATTERED BY A MAGNESIUM  
OXIDE CRYSTAL

ALEXANDER ELLETT AND VICTOR W. COHEN

A beam of neutral potassium atoms is directed to the surface of a MgO crystal, incident at an adjustable angle of from 20 to 80 degrees. Perpendicular to this beam is a slit system which defines a second beam of atoms which have been scattered by the crystal. This second beam is directed to the region between a pair of Stern-Gerlach magnet pole pieces. The intensity of the beam is measured by means of a Langmuir-Taylor surface ionization detector.

By measuring the variation in beam intensity with varying angle of incidence of the primary beam on the crystal while the magnetic field is off, one finds that there is no effect of diffraction or regular reflection of the atoms at the crystal surface. The atoms emerge according to the simple cosine law.

With the application of the magnetic field the beam is broken up into two velocity spectra, in each of which the deflection of an