

1935

A Physical Apparatus for the Determination of the Circulation Time of the Blood in Dogs

E. C. McCracken
Iowa State College

Let us know how access to this document benefits you

Copyright ©1935 Iowa Academy of Science, Inc.

Follow this and additional works at: <https://scholarworks.uni.edu/pias>

Recommended Citation

McCracken, E. C. (1935) "A Physical Apparatus for the Determination of the Circulation Time of the Blood in Dogs," *Proceedings of the Iowa Academy of Science*, 42(1), 149-149.

Available at: <https://scholarworks.uni.edu/pias/vol42/iss1/57>

This Research is brought to you for free and open access by the Iowa Academy of Science at UNI ScholarWorks. It has been accepted for inclusion in Proceedings of the Iowa Academy of Science by an authorized editor of UNI ScholarWorks. For more information, please contact scholarworks@uni.edu.

A PHYSICAL APPARATUS FOR THE DETERMINATION OF THE CIRCULATION TIME OF THE BLOOD IN DOGS

E. C. McCRACKEN

A radical departure from previous apparatus for the determination of blood circulation time is described in detail. The ensemble consists of an ionization chamber (Geiger Counter), a high-voltage (to 2500 volts) power supply, a three-stage amplifying system, and an output circuit for recording the discharges which occur when radiations enter the chamber. A separate electrical circuit is utilized to measure the time interval between the injection of the radioactive material (radium C) in the jugular vein of the dog and its arrival in that part of the femoral artery lying in the adductor canal. The technique of separating radium C from radium emanation is described as is also that of shielding the ionization chamber from extraneous radiations. The record obtained during an experimental test is interpreted as to accuracy and physiological significance and the advantages and disadvantages of the method are outlined.

DEPARTMENT OF PHYSICS,
IOWA STATE COLLEGE,
AMES, IOWA.

THE CIRCULATION TIME OF THE BLOOD IN DOGS BEFORE AND DURING THE DIGESTION OF FOOD

E. C. McCRACKEN

The circulation time of the blood in dogs was investigated by an ensemble of physical apparatus consisting of an ionization chamber, a high-voltage power supply, a three-stage amplifying system, and an output circuit for recording and timing discharges which occur in the ionization chamber when radiations such as gamma and beta rays enter. Radium C was injected into the jugular vein of the dog and its arrival in that portion of the femoral artery lying in the adductor canal was detected by the discharges occurring in the